

Hgb & Hct Panel

Synonyms	H&H
Abbreviation	H & H
Specimen type	Whole blood (EDTA)
Container	Lavender
Volume	3 mL Minimum: 1 mL
Collection requirements	Vacutainer must be at least half full and well-mixed, free of clots.
S/H instructions	Analyze asap for optimal accuracy.
Stability	24 hrs at room temp. 48 hrs refrigerated.
Reference range	HGB 13.5-18.0 g/dl (Male) 11.9-16.0 g/dl (Female) HCT 39-54% (Male) 35-47% (Female)
Routine / STAT T.A.T	Within 4 hours / 1 hour
Section	Hematology
LAB code	02020